



—
your partner
in sensor
technology.

CE - Declaration of Conformity

(According to ISO/IEC 17050-1)

Product(s) Type	From Version:	Measure:	Output Signal:
HTS801-yyy <i>For option AM2 and AM3 see independent DoC yyy order code</i>	21023.3	Humidity Temperature	0-20mA, 4-20mA 0-1V, 0-5V, 0-10V digital RS485 digital TCP/IP



E+E Elektronik Ges.m.b.H.
Langwiesen 7
4209 Engerwitzdorf / Austria

We declare under our sole responsibility that this product(s) (see product table above) corresponds to the following regulations and their subsequent modifications:

Directive Ref.	Directive Area
2014/30/EU	Electromagnetic compatibility
2011/65/EU and 2015/863	RoHS

The products comply with the following standards or standardized documents:

Standard	Year of Ratification
EN 61326-1	2013
EN 61326-2-3	2013
EN IEC 63000	2018

Designed for use in industrial environment. Affixing of the CE marking (for the first time): 2023

Test Report: Conformity_HTS801_01.docx

Modification: First edition

Engerwitzdorf, August 16th, 2023

DI Wolfgang Timelthaler
(General Manager)

Martin Birkbauer
(Development Sensors)